

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

Český metrologický institut  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

**Calibration laboratory locations:**

- |   |  |
|---|--|
| 1. <b>Regional Inspectorate Praha</b>                   | Radiová 1136/3, 102 00 Praha 10 – Hostivař |
| 2. <b>Regional Inspectorate České Budějovice</b>        | U Sirkárny 33/5, 370 04 České Budějovice 4 |
| 3. <b>Regional Inspectorate Plzeň</b>                   | Bendova 539/11, 301 00 Plzeň               |
| 4. <b>Regional Inspectorate Liberec</b>                 | Slunečná 23, 460 01 Liberec                |
| 5. <b>Regional Inspectorate Most</b>                    | Vladislava Vančury 1428/7, 434 01 Most     |
| 6. <b>Regional Inspectorate Pardubice</b>               | Průmyslová 455, 530 03 Pardubice           |
| 7. <b>Regional Inspectorate Brno</b>                    | Okružní 31, 638 00 Brno                    |
| 8. <b>Regional Inspectorate Jihlava</b>                 | Romana Havelky 17, 586 01 Jihlava          |
| 9. <b>Regional Inspectorate Kroměříž</b>                | Kotojedy 73, 767 01 Kroměříž               |
| 10. <b>Regional Inspectorate Opava</b>                  | Gudrichova 41, 746 01 Opava                |
| 11. <b>Regional Inspectorate Olomouc</b>                | I.P. Pavlova 671/141, 779 00 Olomouc       |
| 12. <b>Laboratories for Primary Metrology in Prague</b> | V Botanice 4, 150 72 Praha 5               |
| 13. <b>TESTCOM Praha</b>                                | Hvožďanská 3, 148 00 Praha 4               |





Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit	max.	unit					
8	Test sieves	0.005 mm	to	150 mm			(4L +1.5) µm	Calibration on a 2D optical instrument, using slide gauge or limit gauges	614-MP-C102	6, 7
9	Roundness / balls, rings, cylinders, cones (Maximum diameter 350 mm)	-2 mm	to	+2 mm			Q[0.025;0.01R] µm	Direct measurement by a ring gauge	614-MP-C103	7
	Straightness and parallelity / rings, cylinders, cones, straightness and parallelity standards (Maximum horizontal path: 200 mm Maximum vertical path: 300 mm)	-2 mm	to	+2 mm			Q[0.2;0.01R] µm			
10	Diameter / thread limit gauges, plain limit gauges	1 mm	to	100 mm			2.5 µm	Calibration on IAC MasterScanner XP 10060 analyzer	614-MP-C106	7
11	Gauges, standards, artefacts	0 mm	to	600 mm			(2.5L +1.2) µm	Calibration on a linear height gauge	614-MP-C104	6
		0 mm	to	2,550 mm			Q[0.09; 0.4L] µm	Measurement on a coordinate measuring machine	815-MP-C503	7, 12
12*	Coordinate measuring machine / system  - tactile - optical - multisensor	0 m	to	6 m			Q[0.01;0.2L] µm	Comparison with the standard	815-MP-C501	7, 12
		0 m	to	3 m			Q[0.01;0.2L] µm			
		0 m	to	3 m			Q[0.01;0.2L] µm			



11\_01-P508b K-20240701





Accredited entity according to ČSN EN ISO/IEC 17025:2018:

Český metrologický institut  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

- <sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.
- <sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.
- <sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).
- L nominal length expressed in metres



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

**CMC for the field of measured quantity: Plane angle**

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion		
		min.	unit							
1	Autocollimators	-4,000	"	to	+4,000 "			Comparison with a small angle generator	411-MP-C004	4
2*	Levels, autocollimators, inclinometers	-180	°	to	180 °		0.05"	Comparison with a rotary table or autocollimator	411-MP-C004	4
3*	Index heads and tables, goniometers, instruments for checking angular divisions, angle encoders, torque wrench angle sensors, built-in angle sensors, polygons, angle gauges, optical prisms	0	°	to	360 °		0.03"	Comparison with an optical polygon, autocollimator or rotary table	411-MP-C006	4, 9
4	Polygons, optical prisms, angle gauges	0	°	to	360 °		0.05"	Measurement using two autocollimators	411-MP-C006	4
5	Angle gauges	0	°	to	360 °		60"	Comparison with angle gauges	411-MP-C006	4

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

φ nominal angle of rotation in [ " ]











Accredited entity according to ČSN EN ISO/IEC 17025:2018:

Český metrologický institut  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

CMC for the field of measured quantity: Mass

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min.	unit					
1	Mass of weights	1 mg	to	20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg 50 kg 100 kg 200 kg 500 kg 1,000 kg	0.0010 mg 0.0013 mg 0.0016 mg 0.0020 mg 0.0026 mg 0.003 mg 0.004 mg 0.005 mg 0.006 mg 0.008 mg 0.010 mg 0.016 mg 0.030 mg 0.080 mg 0.15 mg 0.30 mg 0.80 mg 1.5 mg 3.3 mg 8 mg 50 mg 100 mg 250 mg 500 mg	Comparison with a reference weight	612-MP-C131	1 to 11



The Appendix is an integral part of  
Certificate of Accreditation No.: 401/2024 of 15/08/2024

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min.	unit					
2*	Balances with non-automatic and automatic function	0 g	to	20 kg	Weight E <sub>2</sub>	5·10 <sup>-7</sup>	Loading using a reference weight	612-MP-C132
		20 kg	to	50 kg	F <sub>1</sub>	1.6·10 <sup>-6</sup>		1-11
		50 kg	to	600 kg	F <sub>2</sub>	5·10 <sup>-6</sup>		
		600 kg	to	200,000 kg	M	1.6·10 <sup>-5</sup>		

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).





Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min.	unit					
7*	Velocity of transport vehicles and objects	1 km·h <sup>-1</sup>	to	999 km·h <sup>-1</sup>		0.2 km·h <sup>-1</sup>	Comparison with a standard speedometer Measurement using an optoelectronic delay line	812-MP-C209
		0.1 km·h <sup>-1</sup>	to	320 km·h <sup>-1</sup>		0.01 %		
8*	Revolution – impulse generators and meters <sup>5</sup>	0.01 min <sup>-1</sup>	to	100,000 min <sup>-1</sup>		1 %	Contact method	812-MP-C212
		0.01 s <sup>-1</sup>	to	10,000,000 s <sup>-1</sup>		10 <sup>-10</sup>	Contactless method	12
9*	Linear motion velocity / velocity meters for mechanical parts						Measurement of time and distance using an electric, optoelectronic or electronic delay line	812-MP-C201
		0.01 mm·s <sup>-1</sup>	to	1,000 m·s <sup>-1</sup>		0.01 %		12
10*	Linear motion acceleration / acceleration meters for mechanical parts						Measurement of time and distance using an electric, optoelectronic or electronic delay line	812-MP-C201
		-200 m·s <sup>-2</sup>	to	200 m·s <sup>-2</sup>		0.01 %		12

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

<sup>4</sup> Acceleration can be specified also in g, sensor sensitivity in pC/g, resp. mV/g units, where 1 g = 9.81 m·s<sup>-2</sup>

<sup>5</sup> Revolutions - impulses can be specified also in Hz as the number of revolutions - impulses per 1 s.

<sup>6</sup> Acc. To the type of transmission expression



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

**CMC for the field of measured quantity: Force, mechanical tests**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit	max.	unit					
1	Force / force-proving instruments							Calibration using a standard instrument or by suspending weights (ČSN EN ISO 376)	811-MP-C101	12
		1 N 20 kN	to	20 kN 1 MN			0.00004 0.0002			
2*	Force / force-proving instruments							Calibration by a force transmission standard (force-proving instrument) or by suspending weights	811-MP-C111	1, 4, 6, 9, 10
		0.02 N 5 N	to	5 N 2 MN			0.002 0.001			
3*	Torque / torque measuring instruments							Calibration by a standard torque measuring instrument, torque transmission standard (torque sensor) or using weights and torque arms	811-MP-C102	4, 9, 12
		0.02 N·m 0.2 N·m 1 N·m 100 N·m 1 kN·m	to	0.2 N·m 1 N·m 100 N·m 1 kN·m 10 kN·m			0.001 0.0009 0.0008 0.0001 0.0004			
4*	Torque / torque wrenches							Calibration by a torque transmission standard (torque sensor) (ČSN EN ISO 6789-1, ČSN EN ISO 6789-2)	811-MP-C103	4, 9, 12
		0.02 N·m	to	3 kN·m			0.005			



11\_01\_P508b K-20240701

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit	max.	unit					
5*	Torque / equipment for the calibration of torque wrenches	0.02 N·m 500 N·m	to	500 N·m 10 kN·m			0.001 0.0004	Calibration by reference torque wrench, torque transmission standard (torque sensor) or by means of weights and torque arms	811-MP-C104	9, 12
6*	Torque / tightening tools and systems	0.25 N·m	to	10 kN·m			0.005	Calibration by a calibration device for torque wrenches, torque screwdrivers and tighteners	411-MP-C103	4, 9
7*	Machines for mechanical testing of materials	0.02 N 0.02 N 200 kN	to	5 MN 2 MN 2 MN	Pressure Tension Self-setting of the upper pressure plate	0.28 % 0.28 % 0.34 %	Calibration by a force transmission standard (force-proving instrument) or weights (ASTM E4, ČSN EN ISO 7500)	151-MP-C001; 151-MP-C004 ČSN EN 12390-4 Annex A	1, 4, 6, 10 151-MP-C001 1, 4, 6, 10	
8*	Pendulum hammers for notch impact strength testing of materials	0.01 kN/s	to	150 kN/s	Increase in force	0.7 %	Direct comparison with force and time standards, program AED PANEL32	151-MP-C006	6	
		0.01 J	to	2.5 kJ		0.42 % + 0.1 J	Calibration by a force transmission standard (force-proving instrument) and length standard (ČSN EN ISO 148-2,	151-MP-C002	1, 6, 10	





Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
		20 HBW	to	650 HBW	Brinell scale with 2.5mm and 1mm balls	0.24 %	ČSN EN ISO 6506-3, ASTM E10	813-MP-C301
				Scale with a 5mm ball	0.20 %			
				Scale with a 10 mm ball	0.20 %			
		1 ShA		110 ShA	Shore	0.26 Sh	Direct measurement using a standard hardness tester	813-MP-C308
10	Rockwell penetrating bodies - cone angle - radius of curvature	118 °	to	122 °		0.04°	ČSN EN ISO 6508-2	813-MP-C301
		100 µm	to	300 µm		0.08 µm		
11	Vickers penetrating bodies - angle of opposite walls	134 °	to	138 °		0.04°	ČSN EN ISO 6507-2	813-MP-C301
12*	Hardness – hardness testers	10 HRA	to	100 HRA	Rockwell A	0.38 HR	ČSN EN ISO 6508-2, ASTM E18	813-MP-C307
		10 HRBW	to	110 HRBW	Rockwell B			
		10 HRC	to	80 HRC	Rockwell C			
		10 HRD	to	90 HRD	Rockwell D			
		10 HREW	to	110 HREW	Rockwell E			
		10 HRFW	to	100 HRFW	Rockwell F			
		10 HRGW	to	100 HRGW	Rockwell G			
		10 HRHW	to	110 HRHW	Rockwell H			





Accredited entity according to ČSN EN ISO/IEC 17025:2018:

Český metrologický institut  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

- <sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.
- <sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies..
- <sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).









Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

**CMC for the field of measured quantity: Temperature**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit	max.	unit					
1	Glass thermometers					Division < 0.1 °C	0.015 °C 0.020 °C 0.040 °C 0.050 °C	Comparison with a reference resistance temperature sensor in liquid bath.	133-MP-C001	1, 6, 7
		-90 °C	to	100 °C		0.1 °C ≤ division < 1 °C	0.15 °C 0.20 °C			
		100 °C	to	150 °C						
		150 °C	to	210 °C						
		210 °C	to	360 °C						
		360 °C	to	420 °C						
		420 °C	to	550 °C						
2*	Resistance temperature sensors				0.01 °C		0.002 °C	Direct measurement at triple point of water	133-MP-C002	1, 7
					660.32 °C		0.05 °C	Comparison with a reference resistance temperature sensor at a fixed point		
					-196 °C		0.03 °C	Comparison with a reference resistance temperature sensor near the boiling point of nitrogen		
					-100 °C	to	-90 °C	Comparison with a reference resistance temperature sensor in a vertical furnace		
					-90 °C	to	-80 °C	Comparison with a reference resistance temperature sensor in liquid bath.		
					-80 °C	to	160 °C			
					160 °C	to	300 °C			
					300 °C	to	420 °C			



11\_01\_P5086 K-20240701

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion	
		min.	unit	max.	unit						
3	Thermocouple temperature sensors	420 °C	to	550 °C		Noble metals	0.05 °C		133-MP-C003	1, 7	
		550 °C	to	660 °C			0.09 °C	Comparison with a reference resistance temperature sensor in a vertical furnace			
		0 °C	to	220 °C			0.4 °C	Comparison with a reference resistance temperature sensor in liquid bath.			
		220 °C	to	550 °C			0.5 °C				
		550 °C	to	1,100 °C			0.8 °C	Comparison with a reference thermoelectric temperature sensor in a horizontal furnace			
		1,100 °C	to	1,300 °C		Base metals	1.3 °C				
		1,300 °C	to	1,600 °C			2.1 °C				
		-196 °C					0.3 °C	Comparison with a reference resistance temperature sensor near the boiling point of nitrogen			
		-100 °C					0.3 °C	Comparison with a reference resistance temperature sensor in a vertical furnace			
		-90 °C					0.2 °C	Comparison with a reference resistance temperature sensor in liquid bath.			
		220 °C	to	550 °C			0.4 °C				
		550 °C	to	1,100 °C			1.0 °C	Comparison with a reference thermoelectric temperature sensor in a horizontal furnace			
		1,100 °C	to	1,300 °C			1.4 °C				



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit	max.	unit					
4*	Indication thermometers including temperature measuring chains, their temperature probes and characterization of thermal chambers						0.002 °C	Direct measurement at triple point of water	133-MP-C004	1, 6, 7



The Appendix is an integral part of  
Certificate of Accreditation No.: 401/2024 of 15/08/2024

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion	
		min.	unit	max.	unit						
		550 °C	to	660 °C			0.09 °C	Comparison with a reference resistance temperature sensor in a vertical furnace			
		660 °C	to	1,100 °C				0.8 °C 1.3 °C 2.1 °C	Comparison with a reference thermoelectric temperature sensor in a horizontal furnace		
5*	Infrared non-contact thermometers and measuring chains of infrared non-contact thermometers	1,100 °C	to	1,200 °C			Measuring chains with range limited to (-15 to 500) °C			133-MP-C005	
		1,200 °C	to	1,600 °C							
		-30 °C	to	-15 °C							
		-15 °C	to	0 °C							
		0 °C	to	20 °C							
		20 °C	to	100 °C							
		100 °C	to	200 °C							
		200 °C	to	300 °C							
		300 °C	to	400 °C							
		400 °C	to	500 °C							
		500 °C	to	600 °C							
		600 °C	to	700 °C							
6*	Measuring chains	-200 °C	to	0 °C		"K" <sup>4</sup> type Thermoelectric temperature sensor	0.1 % + 0.1 °C 0.007 % + 0.1 °C 0.017 %	Simulation of sensor electrical input signal	133-MP-C006	1	
		0 °C	to	1,000 °C							
		1,000 °C	to	1,372 °C							
		-200 °C	to	0 °C		"J" <sup>4</sup> type Thermoelectric temperature sensor	0.06 % + 0.08 °C				

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
		0 °C	to	1,200 °C		0.006 % + 0.08 °C		1
		-200 °C	to	-100 °C	“N” <sup>4</sup> type Thermoelectric temperature sensor	0.2 %		
		-100 °C	to	0 °C		0.05 % + 0.15 °C		
		0 °C	to	800 °C		0.15 °C		
		800 °C	to	1 300 °C		0.01 % + 0.07 °C		
		0 °C	to	100 °C	“S” <sup>4</sup> type Thermoelectric temperature sensor	0.7 °C		
		100 °C	to	300 °C		0.55 °C		
		300 °C	to	1,768 °C		0.45 °C		
		-200 °C	to	0 °C	“T” type Thermoelectric temperature sensor <sup>4</sup>	0.1 % + 0.1 °C		
		0 °C	to	400 °C		0.1 °C		1
		200 °C	to	500 °C	“B” type Thermoelectric temperature sensor <sup>4</sup>	2 °C		
		500 °C	to	800 °C		0.8 °C		
		800 °C	to	1,820 °C		0.5 °C		
		0 °C	to	150 °C	“R” type Thermoelectric temperature sensor <sup>4</sup>	0.7 °C	1	1
		150 °C	to	400 °C		0.45 °C		
		400 °C	to	1,768 °C		0.4 °C		
		-200 °C	to	0 °C	Pt 100 Resistance temperature sensor	0.05 °C	1	1
		0 °C	to	850 °C		0.014 % + 0.05 °C		
		-200 °C	to	-150 °C	Pt 1000 Resistance temperature sensor <sup>4</sup>	0.011 °C	1	1
		-150 °C	to	-50 °C		0.03 °C		

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

Český metrologický institut  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location		
		min.	unit							
7	Platinum resistance thermometers	-50 °C	to	0 °C		0.043 °C + 0.043		1		
		0 °C	to	850 °C		0.019 % °C				
		-1 V	to	1 V	Voltage output of transducers	0.007 % + 4 µV				
		1 V	to	24 V		0.007 % + 0.1 mV				
8	Thermocouple temperature sensors	0 mA	to	25 mA	Current output of transducers	0.01 % + 1 µA	Direct measurement at a fixed point	112-MP-C001 1		
		25 mA	to	55 mA		0.01 % + 2 µA				
		-189.3442 °C		Temperature scale definition points ITS-90 – triple point Ar	0.9 mK					
9	Contactless thermometers (TC)	1153 °C		Measuring point diameter less than 5 mm	Fe-C	1.2 °C	Comparison with a standard at eutectic fixed point	112-MP-C002 1		
		1324 °C			Co-C	0.9 °C				
		156.5985 °C		In	0.2 °C	Direct measurement at a fixed point	112-MP-C003 1			
		231.928 °C		Sn	0.2 °C					
		660.323 °C		Al	0.15 °C	Comparison with a reference standard				
		1,084.62 °C		Cu	0.10 °C					
		100 °C	to	300 °C	0.5 °C					
		300 °C	to	600 °C	0.7 °C					
		600 °C	to	1,000 °C	0.8 °C					
		1,000 °C	to	1,300 °C	1.0 °C					
		1,300 °C	to	1,800 °C	1.5 °C					



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
 CAB number 2202, CMI Calibration Laboratory  
 Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion		
		min.	unit	max.	unit							
10*	Black bodies	-30 °C	to	45 °C			0.2 °C	Direct comparison of two black bodies using a transfer thermometer	112-MP-C004	1		
		45 °C	to	230 °C								
		230 °C	to	600 °C								
		600 °C	to	1,000 °C			0.5 °C	Comparison with a reference standard				
		1,000 °C	to	1,300 °C			0.6 °C					
		1,300 °C	to	1,800 °C			1.0 °C					
11*	Thermal cameras	-30 °C	to	500 °C			0.8 °C	Comparison with the standard	112-MP-C005	1, 7		
		500 °C	to	1,000 °C			1.0 °C					
		1,000 °C	to	1,300 °C			1.5 °C					
		1,300 °C	to	1,800 °C			2.0 °C					
12	Temperature calibrators	-50 °C	to	1,760 °C			Comparison with a calibrator or electrical measurement with a multimeter	611-MP-C130	1, 7			
		-50 °C	to	1,760 °C								
		0 °C	to	1,820 °C								
		-210 °C	to	1,200 °C								
		-270 °C	to	400 °C								
		-270 °C	to	1,000 °C								
		-270 °C	to	1,370 °C								
		-270 °C	to	1,300 °C								
		-200 °C	to	100 °C								
		-200 °C	to	800 °C								
		0 °C	to	2,500 °C								
		-200 °C	to	900 °C								



11\_01-P508b K-20240701

-3-

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit	max.	unit					
		-200 °C	to	850 °C		RTD - Pt100	0.02 °C			
		-200 °C	to	850 °C		RTD - Pt200	0.02 °C			
		-200 °C	to	850 °C		RTD - Pt500	0.03 °C			
		-200 °C	to	850 °C		RTD - Pt1000	0.02 °C			
		-60 °C	to	250 °C		RTD - Ni100	0.01 °C			
		-60 °C	to	250 °C		RTD - Ni1000	0.01 °C			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

<sup>4</sup> Applies to the CJC temperature = 0 °C.



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

**CMC for the field of measured quantity: Air humidity**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min.	unit					
1*	Relative humidity / hygrometers and humidity measuring chains including humidity probes, characterization of climatic chambers	5 % RH	to	30 % RH	Air temperature (10 to 90) °C	0.6 % RH	Comparison with a standard hygrometer	636-MP-C119
		30 % RH	to	50 % RH		0.7 % RH		
		50 % RH	to	70 % RH		0.8 % RH		
		70 % RH	to	80 % RH		0.9 % RH		
		80 % RH	to	90 % RH		1.0 % RH		
		90 % RH	to	95 % RH		1.5 % RH		
2	Dew point temperature / hygrometers	-75 °C	to	-65 °C		0.24 °C	Comparison with a standard thermometer	636-MP-C120
		-65 °C	to	-50 °C		0.16 °C		
		-50 °C	to	-30 °C		0.10 °C		
		-30 °C	to	60 °C		0.08 °C		
		60 °C	to	80 °C		0.10 °C		
		80 °C	to	90 °C		0.15 °C		

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

**CMC for the field of measured quantity: Electrical quantities**

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
1*	DC voltage / DC voltage sources and meters							
		0 mV	to	20 mV	$15 \cdot 10^{-6} \cdot U + 0.05 \mu V$			
		20 mV	to	200 mV	$5.0 \cdot 10^{-6} \cdot U$			
		200 mV	to	2 V	$2.6 \cdot 10^{-6} \cdot U$			
		2 V	to	20 V	$2.1 \cdot 10^{-6} \cdot U$			
		10 V	to	10 V	$1.5 \mu V$			
		20 V	to	1,100 V	$2.9 \cdot 10^{-6} \cdot U$			
2	DC voltage / DC voltage sources and meters							
		0 V	to	2 V	$1.5 \cdot 10^{-6} \cdot U + 0.05 \mu V$			
		2 V	to	20 V	$0.8 \cdot 10^{-6} \cdot U$			
		20 V	to	1,100 V	$1.6 \cdot 10^{-6} \cdot U$			
3*	AC voltage / AC voltage sources and meters							
		0.9 mV	to	2 mV	10 Hz to 75 kHz 75 kHz to 400 kHz 400 kHz to 750 kHz 750 kHz to 1 MHz	0.20 % 0.31 % 0.33 % 0.35 %		
		2 mV	to	20 mV	10 Hz to 25 kHz 25 kHz to 75 kHz 75 kHz to 200 kHz 200 kHz to 400 kHz 400 kHz to 750 kHz 750 kHz to 1 MHz	$360 \cdot 10^{-6} \cdot U$ $330 \cdot 10^{-6} \cdot U$ $530 \cdot 10^{-6} \cdot U$ $710 \cdot 10^{-6} \cdot U$ $790 \cdot 10^{-6} \cdot U$ $1100 \cdot 10^{-6} \cdot U$		



11\_01-P508b K-20240701

-3-

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
		20 mV	to	200 mV	10 Hz to 25 kHz 25 kHz to 75 kHz 75 kHz to 200 kHz 200 kHz to 400 kHz 400 kHz to 750 kHz 750 kHz to 1 MHz	98·10 <sup>-6</sup> ·U 190·10 <sup>-6</sup> ·U 370·10 <sup>-6</sup> ·U 590·10 <sup>-6</sup> ·U 650·10 <sup>-6</sup> ·U 1,000·10 <sup>-6</sup> ·U		
		200 mV	to	2 V	10 Hz to 35 Hz 35 Hz to 40 kHz 40 kHz to 75 kHz 75 kHz to 200 kHz 200 kHz to 400 kHz 400 kHz to 750 kHz 750 kHz to 1 MHz	37·10 <sup>-6</sup> ·U 26·10 <sup>-6</sup> ·U 37·10 <sup>-6</sup> ·U 48·10 <sup>-6</sup> ·U 120·10 <sup>-6</sup> ·U 230·10 <sup>-6</sup> ·U 590·10 <sup>-6</sup> ·U		
		2 V	to	20 V	10 Hz to 35 Hz 35 Hz to 40 kHz 40 kHz to 75 kHz 75 kHz to 200 kHz 200 kHz to 400 kHz 400 kHz to 750 kHz 750 kHz to 1 MHz	37·10 <sup>-6</sup> ·U 26·10 <sup>-6</sup> ·U 33·10 <sup>-6</sup> ·U 40·10 <sup>-6</sup> ·U 110·10 <sup>-6</sup> ·U 210·10 <sup>-6</sup> ·U 560·10 <sup>-6</sup> ·U		
		20 V	to	200 V	10 Hz to 35 Hz 35 Hz to 175 Hz 175 Hz to 40 kHz 40 kHz to 75 kHz 75 kHz to 150 kHz 150 kHz to 200 kHz	42·10 <sup>-6</sup> ·U 37·10 <sup>-6</sup> ·U 28·10 <sup>-6</sup> ·U 40·10 <sup>-6</sup> ·U 71·10 <sup>-6</sup> ·U 240·10 <sup>-6</sup> ·U		



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion	
		min.	unit						
		200 V	to	1,100 V	10 Hz to 5 kHz 5 kHz to 25 kHz 25 kHz to 40 kHz 40 kHz to 750 kHz Generation only to 750 V	40·10 <sup>-6</sup> ·U 45·10 <sup>-6</sup> ·U 76·10 <sup>-6</sup> ·U 120·10 <sup>-6</sup> ·U 350·10 <sup>-6</sup> ·U			
4	Low AC voltage / AC voltage meters	0.1 µV to 1 µV		50 Hz to 200 Hz 200 Hz to 10 kHz 10 kHz to 100 kHz	0.50 % 0.20 % 0.70 %	Comparison or measurement using an impedance divider	611-MP-C061	7	
		1 µV to 1 mV		50 Hz to 200 Hz 200 Hz to 10 kHz 10 kHz to 100 kHz	0.20 % 0.07 % 0.30 %				
		0 pA to 1 pA 1 pA to 20 pA 20 pA to 200 pA 0.2 nA to 2 nA 2 nA to 20 nA 20 nA to 200 nA 0.2 µA to 2 µA				1 fA 0.13 % 0.11 % 0.08 % 0.06 % 0.05 % 0.04 %	Direct measurement with a picoammeter	611-MP-C034	7
		DC current / Low DC current generators							
5	DC current / Low DC current meters	0 pA to 1 pA 1 pA to 20 pA 20 pA to 200 pA							
		0 pA to 1 pA 1 pA to 20 pA 20 pA to 200 pA							



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
6*	DC current / DC current generators	0.2 nA	to	2 nA	0.03 %	Direct measurement with a multimeter or indirect measurement with a current shunt	611-MP-C097	1, 7
		2 nA	to	20 nA	0.025 %			
		20 nA	to	200 nA	0.020 %			
		0.2 µA	to	2 µA	0.008 %			
		0 µA	to	1 µA	0.10 nA			
	DC current / DC current meters	1 µA	to	100 µA	$21 \cdot 10^{-6} \cdot I$	Direct generation with a calibrator	611-MP-C098	1, 7
		100 µA	to	200 µA	$12 \cdot 10^{-6} \cdot I$			
		200 µA	to	2 A	$6.0 \cdot 10^{-6} \cdot I$			
		2 A	to	20 A	$15 \cdot 10^{-6} \cdot I$			
		20 A	to	100 A	$20 \cdot 10^{-6} \cdot I$			
7*	AC current / AC current generators	0 µA	to	1 µA	0.10 nA	Direct measurement with a multimeter or indirect measurement with current shunts	611-MP-C097	1, 7
		1 µA	to	100 µA	$21 \cdot 10^{-6} \cdot I$			
		100 µA	to	200 µA	$15 \cdot 10^{-6} \cdot I$			
		200 µA	to	20 mA	$6.0 \cdot 10^{-6} \cdot I$			
		20 mA	to	200 mA	$8.0 \cdot 10^{-6} \cdot I$			



-3- 11\_01-P508b K-20240701

The Appendix is an integral part of  
Certificate of Accreditation No.: 401/2024 of 15/08/2024

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
				10 kHz to 30 kHz	$300 \cdot 10^{-6} \cdot I$			
		2 mA	to	20 mA	10 Hz to 20 Hz	$90 \cdot 10^{-6} \cdot I$		
					20 Hz to 40 Hz	$75 \cdot 10^{-6} \cdot I$		
					40 Hz to 1 kHz	$65 \cdot 10^{-6} \cdot I$		
					1 kHz to 10 kHz	$70 \cdot 10^{-6} \cdot I$		
					10 kHz to 30 kHz	$300 \cdot 10^{-6} \cdot I$		
		20 mA	to	200 mA	10 Hz to 20 Hz	$80 \cdot 10^{-6} \cdot I$		
					20 Hz to 1 kHz	$65 \cdot 10^{-6} \cdot I$		
					1 kHz to 5 kHz	$70 \cdot 10^{-6} \cdot I$		
					5 kHz to 10 kHz	$80 \cdot 10^{-6} \cdot I$		
					10 kHz to 30 kHz	$300 \cdot 10^{-6} \cdot I$		
		200 mA	to	2 A	10 Hz to 20 Hz	$120 \cdot 10^{-6} \cdot I$		
					20 Hz to 40 Hz	$100 \cdot 10^{-6} \cdot I$		
					40 Hz to 1 kHz	$90 \cdot 10^{-6} \cdot I$		
					1 kHz to 5 kHz	$100 \cdot 10^{-6} \cdot I$		
					5 kHz to 10 kHz	$120 \cdot 10^{-6} \cdot I$		
					10 kHz to 30 kHz	$500 \cdot 10^{-6} \cdot I$		
		2 A	to	20 A	10 Hz to 20 Hz	$180 \cdot 10^{-6} \cdot I$		
					20 Hz to 40 Hz	$160 \cdot 10^{-6} \cdot I$		
					40 Hz to 1 kHz	$110 \cdot 10^{-6} \cdot I$		
					1 kHz to 5 kHz	$140 \cdot 10^{-6} \cdot I$		
					5 kHz to 10 kHz	$160 \cdot 10^{-6} \cdot I$		
		20 A	to	100 A	10 Hz to 20 Hz	$250 \cdot 10^{-6} \cdot I$		
					20 Hz to 40 Hz	$160 \cdot 10^{-6} \cdot I$		
					40 Hz to 1 kHz	$120 \cdot 10^{-6} \cdot I$		
					1 kHz to 5 kHz	$150 \cdot 10^{-6} \cdot I$		
					5 kHz to 10 kHz	$200 \cdot 10^{-6} \cdot I$		



-3- 11\_01\_P508b K-20240701

The Appendix is an integral part of  
Certificate of Accreditation No.: 401/2024 of 15/08/2024

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
AC current / AC current meters	AC current / AC current meters	9	µA	to	200	µA	10 Hz to 35 Hz 35 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 10 kHz to 30 kHz	130·10 <sup>-6</sup> ·I 100·10 <sup>-6</sup> ·I 160·10 <sup>-6</sup> ·I 800·10 <sup>-6</sup> ·I 2,000·10 <sup>-6</sup> ·I
		200	µA	to	2	mA	10 Hz to 35 Hz 35 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 30 kHz	120·10 <sup>-6</sup> ·I 90·10 <sup>-6</sup> ·I 140·10 <sup>-6</sup> ·I 700·10 <sup>-6</sup> ·I
		2	mA	to	200	mA	10 Hz to 35 Hz 35 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 10 kHz to 30 kHz	120·10 <sup>-6</sup> ·I 75·10 <sup>-6</sup> ·I 130·10 <sup>-6</sup> ·I 140·10 <sup>-6</sup> ·I 700·10 <sup>-6</sup> ·I
		200	mA	to	2	A	10 Hz to 35 Hz 35 Hz to 1 kHz 1 kHz to 10 kHz	170·10 <sup>-6</sup> ·I 120·10 <sup>-6</sup> ·I 220·10 <sup>-6</sup> ·I
		2	A	to	20	A	10 Hz to 35 Hz 35 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	310·10 <sup>-6</sup> ·I 200·10 <sup>-6</sup> ·I 300·10 <sup>-6</sup> ·I 690·10 <sup>-6</sup> ·I
		20	A	to	100	A	10 Hz to 35 Hz 35 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	350·10 <sup>-6</sup> ·I 200·10 <sup>-6</sup> ·I 300·10 <sup>-6</sup> ·I 950·10 <sup>-6</sup> ·I



11 01-P508b K-20240701

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
8	DC resistance / DC resistance standards and electrical resistance meters					Comparison by the ratio method – measurement of voltage drops at constant measuring current or direct measurement of reference resistance	131-MP-C003, 131-MP-C005, 611-MP-C097, 611-MP-C098	1, 7
		100 µΩ			0.72 nΩ			
		1 mΩ			3.5 nΩ			
		10 mΩ			26 nΩ			
		20 mΩ			44 nΩ			
		100 mΩ			86 nΩ			
		1 Ω			0.76 µΩ			
		10 Ω			7.1 µΩ			
		25 Ω			16 µΩ			
		100 Ω			35 µΩ			
		1 kΩ			0.35 mΩ			
		10 kΩ			3.8 mΩ			
		100 kΩ			59 mΩ			
		1 MΩ			4.0 Ω			
		10 MΩ			70 Ω			
		100 MΩ			160 Ω			
		1 GΩ			100 kΩ			
		10 GΩ			5.0 MΩ			
		100 GΩ			25 MΩ			
		1 TΩ			200 MΩ			



11\_01-P508b K-20240701

-3-

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
9*	DC resistance / DC resistance standards and electrical resistance meters							
		0 $\mu\Omega$	to	100 $\mu\Omega$	$3.5 \cdot 10^{-6} \cdot R + 0.5 \text{ n}\Omega$			
		100 $\mu\Omega$	to	200 $\mu\Omega$	$7.2 \cdot 10^{-6} \cdot R$			
		200 $\mu\Omega$	to	1 m $\Omega$	$6.0 \cdot 10^{-6} \cdot R$			
		1 m $\Omega$	to	2 m $\Omega$	$3.5 \cdot 10^{-6} \cdot R$			
		2 m $\Omega$	to	10 m $\Omega$	$5.6 \cdot 10^{-6} \cdot R$			
		10 m $\Omega$	to	20 m $\Omega$	$2.6 \cdot 10^{-6} \cdot R$			
		20 m $\Omega$	to	100 m $\Omega$	$1.6 \cdot 10^{-6} \cdot R$			
		100 m $\Omega$	to	200 m $\Omega$	$0.86 \cdot 10^{-6} \cdot R$			
		200 m $\Omega$	to	1 $\Omega$	$1.6 \cdot 10^{-6} \cdot R$			
		1 $\Omega$	to	2 $\Omega$	$0.73 \cdot 10^{-6} \cdot R$			
		2 $\Omega$	to	10 $\Omega$	$1.6 \cdot 10^{-6} \cdot R$			
		10 $\Omega$	to	20 $\Omega$	$0.71 \cdot 10^{-6} \cdot R$			
		20 $\Omega$	to	100 $\Omega$	$1.2 \cdot 10^{-6} \cdot R$			
		100 $\Omega$	to	200 $\Omega$	$0.35 \cdot 10^{-6} \cdot R$			
		200 $\Omega$	to	1 k $\Omega$	$0.93 \cdot 10^{-6} \cdot R$			
		1 k $\Omega$	to	2 k $\Omega$	$0.35 \cdot 10^{-6} \cdot R$			
		2 k $\Omega$	to	10 k $\Omega$	$1.2 \cdot 10^{-6} \cdot R$			
		10 k $\Omega$	to	20 k $\Omega$	$0.38 \cdot 10^{-6} \cdot R$			
		20 k $\Omega$	to	100 k $\Omega$	$0.86 \cdot 10^{-6} \cdot R$			
		100 k $\Omega$	to	200 k $\Omega$	$0.59 \cdot 10^{-6} \cdot R$			
		200 k $\Omega$	to	1 M $\Omega$	$4.1 \cdot 10^{-6} \cdot R$			



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
		1 MΩ	to	2 MΩ	4.0·10 <sup>-6</sup> ·R			
		2 MΩ	to	10 MΩ		7.4·10 <sup>-6</sup> ·R		
		10 MΩ	to	20 MΩ		7.0·10 <sup>-6</sup> ·R		
		20 MΩ	to	200 MΩ		16·10 <sup>-6</sup> ·R		
		200 MΩ	to	1 GΩ		520·10 <sup>-6</sup> ·R		
		1 GΩ	to	2 GΩ		120·10 <sup>-6</sup> ·R		
		2 GΩ	to	10 GΩ		0.52 %		
		10 GΩ	to	20 GΩ		0.12 %		
10	Resistance ratio / Thermometer bridges 0 Hz to 400 Hz	0	to	4	1.6·10 <sup>-8</sup>	Measurement with a set of reference resistors or a simulated resistance ratio generated by an inductive divider	611-MP-C045	1, 7
11	DC Power / DC Power Meters for 1 V to 1,000 V and 1 mA to 120 A	1 mW	to	120 kW		0.0025%		



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion		
		min.	unit							
12	Single-phase and three-phase AC active power / Active electrical power meters for voltage 1 V to 1,000 V, current 1 mA to 200 A, power factor 0 to 1, frequency 15 Hz to 1,000 Hz	1 mW	to	600 kW	25 µW/VA	Measurement by a digital sampling wattmeter	611-MP-C042	7		
13	Single-phase and three-phase AC reactive power / Reactive electrical power meters for voltage 1 V to 1,000 V, current 1 mA to 200 A, power factor 0 to 1, frequency 15 Hz to 1,000 Hz	1 mvar	to	600 kvar	25 µvar/VA	Measurement by a digital sampling wattmeter	611-MP-C042	7		
14	Phase angle / Phase angle meters	0 °	to	360 °	0.1 µV to 1 µV 1 µV to 100 µV 0.1 mV to 1 mV 1 mV to 10 mV  10 mV to 560 V 10 mV to 560 V 10 mV to 100 V	400 Hz to 1.592 kHz  1 Hz to 6 kHz 6 kHz to 50 kHz 50 kHz to 100 kHz	0.10° 0.050° 0.10° 0.050°  0.0010° 0.0020° 0.050°	Measurement on a phase calibrator and impedance divider	611-MP-C061	7
							Digital signal sampling	611-MP-C060		



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand		Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit	max.	unit				
				10 mV to 1 V	100 kHz to 1 MHz	0.050°			
				10 mV to 1 V	1 MHz to 10 MHz	0.10°			
15	Electrical energy single-phase and three- phase / DC and AC power meters for voltage 1 V to 1,000 V, current 1 mA to 120 A, power factor 0 to 1, frequency 0 Hz and 15 Hz to 800 Hz, time 1 s to 3,600 s	1 Ws	to	1296 MWs		0.010 %	Comparison with a reference energy meter Measurement by a digital sampling wattmeter	611-MP-C042	7
16	Flicker / Flicker meters	0.5 Pst	to	10 Pst	50 Hz	0.003 · Pst	Using digital sampling of a signal	611-MP-C043	7
17	pH / pH meters	0 pH	to	14 pH		0.001 pH	Comparison with a calibrator (electric method)	611-MP-C129	7
		-1,000 mV	to	1,000 mV		0.010 mV			
18	Capacity / Electrical capacity meters	10 pF		100 pF	1 kHz	5 aF	Comparison with a standard using impedance bridge	611-MP-C030, 611-MP-C041	7
		1 pF	to	10 pF	1 kHz 200 Hz to 1 kHz 1 kHz 2 kHz to 10 kHz 10 kHz to 1 MHz	50 aF 0.015 % 0.0070 % 0.0015 % 0.0050 % 0.010 %			



11\_01-P508b K-20240701  
-3-

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
				1 MHz to 10 MHz	0.20 %			
		10 pF	to	1 nF	50 Hz to 1 kHz	0.0050 %		
					1 kHz	0.0007 %		
					1 kHz to 10 kHz	0.0050 %		
					10 kHz to 1 MHz	0.010 %		
					1 MHz to 10 MHz	0.20 %		
		1 nF	to	10 nF	50 Hz to 1 kHz	0.0070 %		
					1 kHz	0.0010 %		
					1 kHz to 10 kHz	0.0050 %		
					10 kHz to 1 MHz	0.010 %		
					1 MHz to 10 MHz	0.20 %		
		10 nF	to	100 nF	50 Hz to 200 Hz	0.010 %		
					200 Hz to 1 kHz	0.0030 %		
					1 kHz to 10 kHz	0.0070 %		
					10 kHz to 100 kHz	0.010 %		
					100 kHz to 1 MHz	0.030 %		
		100 nF	to	1 µF	50 Hz to 200 Hz	0.0070 %		
					200 Hz to 10 kHz	0.0050 %		
					10 kHz to 100 kHz	0.015 %		
					100 kHz to 1 MHz	0.050 %		
		1 µF	to	10 µF	20 Hz to 50 Hz	0.010 %		
					50 Hz to 1 kHz	0.0070 %		
					1 kHz	0.0050 %		
					1 kHz to 10 kHz	0.0060 %		
					10 kHz to 20 kHz	0.015 %		
					20 kHz to 100 kHz	0.050 %		
		10 µF	to	100 µF	20 Hz to 50 Hz	0.010 %		



-3+ 01-P508b K-20240701

The Appendix is an integral part of  
Certificate of Accreditation No.: 401/2024 of 15/08/2024

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion		
		min.	unit							
		100 µF	to	1 mF	50 Hz to 1 kHz	0.0050 %				
					1 kHz to 10 kHz	0.0060 %				
					10 kHz to 20 kHz	0.020 %				
		1 mF	to	3 mF	20 Hz to 200 Hz	0.010 %				
					200 Hz to 1 kHz	0.015 %				
					1 kHz to 5 kHz	0.020 %				
19	Loss factor D / Loss factor meters	-0.001	to	0.001	20 Hz to 200 Hz	0.020 %	Comparison with a standard using impedance bridge	611-MP-C030 611-MP-C041		
					200 Hz to 1 kHz	0.025 %				
					0.1 Hz	0.10 %				
					1 kHz	$57 \cdot 10^{-6}$ (abs.)				
					1 kHz to 5 kHz					
					5 kHz to 10 kHz					
					10 kHz to 20 kHz					
					20 kHz to 100 kHz					
					100 kHz to 200 kHz					
					200 kHz to 500 kHz					
					500 kHz to 1 MHz					
					10 pF to 100 pF	13 · 10 <sup>-6</sup> (abs.)				
					50 Hz to 200 Hz	4 · 10 <sup>-6</sup> (abs.)				
					200 Hz to 1 kHz	3 · 10 <sup>-6</sup> (abs.)				
					1 kHz	9 · 10 <sup>-6</sup> (abs.)				
					1 kHz to 5 kHz					



11\_01-P508b K-20240701

-3-

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
				5 kHz to 10 kHz	$18 \cdot 10^{-6}$ (abs.)			
				10 kHz to 20 kHz	$39 \cdot 10^{-6}$ (abs.)			
				20 kHz to 100 kHz	$70 \cdot 10^{-6}$ (abs.)			
				100 kHz to 200 kHz	$150 \cdot 10^{-6}$ (abs.)			
				200 kHz to 500 kHz	$350 \cdot 10^{-6}$ (abs.)			
				500 kHz to 1 MHz	$700 \cdot 10^{-6}$ (abs.)			
				100 pF to 1,000 pF	50 Hz to 200 Hz 200 Hz to 1 kHz 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz	$9 \cdot 10^{-6}$ (abs.) $4 \cdot 10^{-6}$ (abs.) $3 \cdot 10^{-6}$ (abs.) $7 \cdot 10^{-6}$ (abs.) $12 \cdot 10^{-6}$ (abs.) $23 \cdot 10^{-6}$ (abs.) $70 \cdot 10^{-6}$ (abs.) $150 \cdot 10^{-6}$ (abs.) $350 \cdot 10^{-6}$ (abs.) $700 \cdot 10^{-6}$ (abs.)		
				1 nF to 10 nF	50 Hz to 200 Hz 200 Hz to 1 kHz 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz	$79 \cdot 10^{-6}$ (abs.) $7 \cdot 10^{-6}$ (abs.) $9 \cdot 10^{-6}$ (abs.) $24 \cdot 10^{-6}$ (abs.) $41 \cdot 10^{-6}$ (abs.) $79 \cdot 10^{-6}$ (abs.) $70 \cdot 10^{-6}$ (abs.) $150 \cdot 10^{-6}$ (abs.) $350 \cdot 10^{-6}$ (abs.) $700 \cdot 10^{-6}$ (abs.)		



14-01-P508b K-20240701

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
		0.001	to	0.01	1 pF to 10 pF 50 Hz to 200 Hz 200 Hz to 1 kHz 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 10 kHz to 20 kHz	$61 \cdot 10^{-6}$ (abs.) $18 \cdot 10^{-6}$ (abs.) $5 \cdot 10^{-6}$ (abs.) $26 \cdot 10^{-6}$ (abs.) $50 \cdot 10^{-6}$ (abs.) $130 \cdot 10^{-6}$ (abs.)		
				10 pF to 100 pF 50 Hz to 200 Hz 200 Hz to 1 kHz 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 10 kHz to 20 kHz	$20 \cdot 10^{-6}$ (abs.) $13 \cdot 10^{-6}$ (abs.) $5 \cdot 10^{-6}$ (abs.) $15 \cdot 10^{-6}$ (abs.) $23 \cdot 10^{-6}$ (abs.) $43 \cdot 10^{-6}$ (abs.)			
				100 pF to 1,000 pF 50 Hz to 200 Hz 200 Hz to 1 kHz 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 10 kHz to 20 kHz	$12 \cdot 10^{-6}$ (abs.) $7 \cdot 10^{-6}$ (abs.) $5 \cdot 10^{-6}$ (abs.) $9 \cdot 10^{-6}$ (abs.) $14 \cdot 10^{-6}$ (abs.) $25 \cdot 10^{-6}$ (abs.)			
				1 nF to 10 nF 50 Hz to 200 Hz 200 Hz to 1 kHz 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 10 kHz to 20 kHz	$82 \cdot 10^{-6}$ (abs.) $7 \cdot 10^{-6}$ (abs.) $9 \cdot 10^{-6}$ (abs.) $24 \cdot 10^{-6}$ (abs.) $41 \cdot 10^{-6}$ (abs.) $79 \cdot 10^{-6}$ (abs.)			
		0.01	to	0.1	1 pF to 10 pF 50 Hz to 200 Hz 200 Hz to 1 kHz 1 kHz	$150 \cdot 10^{-6}$ (abs.) $120 \cdot 10^{-6}$ (abs.) $30 \cdot 10^{-6}$ (abs.)		



11\_01\_P508b K-20240701

-3-

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
				1 kHz to 5 kHz	$110 \cdot 10^{-6}$ (abs.)			
				5 kHz to 10 kHz	$130 \cdot 10^{-6}$ (abs.)			
				10 pF to 100 pF	50 Hz to 200 Hz	$61 \cdot 10^{-6}$ (abs.)		
					200 Hz to 1 kHz	$52 \cdot 10^{-6}$ (abs.)		
					1 kHz	$30 \cdot 10^{-6}$ (abs.)		
					1 kHz to 5 kHz	$48 \cdot 10^{-6}$ (abs.)		
					5 kHz to 10 kHz	$58 \cdot 10^{-6}$ (abs.)		
				100 pF to 1,000 pF	50 Hz to 200 Hz	$140 \cdot 10^{-6}$ (abs.)		
					200 Hz to 1 kHz	$74 \cdot 10^{-6}$ (abs.)		
					1 kHz	$30 \cdot 10^{-6}$ (abs.)		
					1 kHz to 5 kHz	$70 \cdot 10^{-6}$ (abs.)		
					5 kHz to 10 kHz	$61 \cdot 10^{-6}$ (abs.)		
				1 nF to 10 nF	50 Hz to 200 Hz	$100 \cdot 10^{-6}$ (abs.)		
					200 Hz to 1 kHz	$7 \cdot 10^{-6}$ (abs.)		
					1 kHz	$9 \cdot 10^{-6}$ (abs.)		
					1 kHz to 5 kHz	$24 \cdot 10^{-6}$ (abs.)		
					5 kHz to 10 kHz	$41 \cdot 10^{-6}$ (abs.)		
					10 kHz to 20 kHz	$79 \cdot 10^{-6}$ (abs.)		
		0.1	to	1	1 pF to 1,000 pF	1 kHz	$300 \cdot 10^{-6}$ (abs.)	
					1 nF to 10 nF	50 Hz to 200 Hz	$100 \cdot 10^{-6}$ (abs.)	
						200 Hz to 1 kHz	$7 \cdot 10^{-6}$ (abs.)	
						1 kHz	$9 \cdot 10^{-6}$ (abs.)	
						1 kHz to 5 kHz	$24 \cdot 10^{-6}$ (abs.)	
						5 kHz to 10 kHz	$41 \cdot 10^{-6}$ (abs.)	
						10 kHz to 20 kHz	$79 \cdot 10^{-6}$ (abs.)	



11\_01-P508b K-20240701

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion	
		min.	unit						
		-1	to	1	10 nF to 100 nF	50 Hz to 200 Hz 200 Hz to 1 kHz 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz	$30 \cdot 10^{-6}$ (abs.) $15 \cdot 10^{-6}$ (abs.) $15 \cdot 10^{-6}$ (abs.) $21 \cdot 10^{-6}$ (abs.) $30 \cdot 10^{-6}$ (abs.) $67 \cdot 10^{-6}$ (abs.) $70 \cdot 10^{-6}$ (abs.) $150 \cdot 10^{-6}$ (abs.) $350 \cdot 10^{-6}$ (abs.) $700 \cdot 10^{-6}$ (abs.)		
				100 nF to 1,000 nF	20 Hz to 50 Hz 50 Hz to 5 kHz 5 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz	$100 \cdot 10^{-6}$ (abs.) $30 \cdot 10^{-6}$ (abs.) $60 \cdot 10^{-6}$ (abs.) $80 \cdot 10^{-6}$ (abs.) $90 \cdot 10^{-6}$ (abs.) $200 \cdot 10^{-6}$ (abs.) $450 \cdot 10^{-6}$ (abs.) $900 \cdot 10^{-6}$ (abs.)			
				1 µF to 10 µF	20 Hz to 50 Hz 50 Hz to 5 kHz 5 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 100 kHz	$100 \cdot 10^{-6}$ (abs.) $30 \cdot 10^{-6}$ (abs.) $60 \cdot 10^{-6}$ (abs.) $80 \cdot 10^{-6}$ (abs.) $120 \cdot 10^{-6}$ (abs.)			
				10 µF to 100 µF	20 Hz to 50 Hz 50 Hz to 200 Hz 200 Hz to 5 kHz 5 kHz to 10 kHz	$100 \cdot 10^{-6}$ (abs.) $30 \cdot 10^{-6}$ (abs.) $50 \cdot 10^{-6}$ (abs.) $100 \cdot 10^{-6}$ (abs.)			



11\_01-P508b K-20240701

-3-

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion			
		min.	unit								
20	AC resistance / AC resistance meters			10 kHz to 20 kHz	$220 \cdot 10^{-6}$ (abs.)						
				20 kHz to 100 kHz	$500 \cdot 10^{-6}$ (abs.)						
				100 µF to 1,000 µF	20 Hz to 50 Hz	$100 \cdot 10^{-6}$ (abs.)					
					50 Hz to 1 kHz	$50 \cdot 10^{-6}$ (abs.)					
					1 kHz to 5 kHz	$100 \cdot 10^{-6}$ (abs.)					
					5 kHz to 10 kHz	$160 \cdot 10^{-6}$ (abs.)					
					10 kHz to 20 kHz	$350 \cdot 10^{-6}$ (abs.)					
				1 mF to 3 mF	20 Hz to 50 Hz	$100 \cdot 10^{-6}$ (abs.)					
					50 Hz to 200 Hz	$50 \cdot 10^{-6}$ (abs.)					
					200 Hz to 5 kHz	$100 \cdot 10^{-6}$ (abs.)					
		3 mF to 10 mF		20 Hz to 1 kHz	$100 \cdot 10^{-6}$ (abs.)						
20	AC resistance / AC resistance meters	1 mΩ to 10 mΩ	1 mΩ to 10 mΩ	20 Hz to 50 Hz	0.015 %	Comparison with a standard using impedance bridge	611-MP-C040, 611-MP-C041	7			
		10 mΩ to 100 mΩ		50 Hz to 20 kHz	0.010 %						
		100 mΩ to 1 Ω		20 kHz to 100 kHz	0.015 %						
		1 Ω to 10 kΩ		100 kHz to 1 MHz	0.10 %						



11\_01-P5086 K-20240701

-3-

The Appendix is an integral part of  
Certificate of Accreditation No.: 401/2024 of 15/08/2024

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion	
		min.	unit						
				200 kHz to 500 kHz	0,010 %				
				500 kHz to 1 MHz	0,015 %				
				1 MHz to 10 MHz	0,20 %				
		10 kΩ to 100 kΩ		20 Hz to 50 Hz	0,0070 %				
				50 Hz to 20 kHz	0,0050 %				
				20 kHz to 100 kHz	0,0070 %				
				100 kHz to 200 kHz	0,010 %				
				200 kHz to 500 kHz	0,015 %				
				500 kHz to 1 MHz	0,025 %				
				1 MHz to 10 MHz	5,0 %				
		100 kΩ to 1 MΩ		20 Hz to 50 Hz	0,015 %				
				50 Hz to 20 kHz	0,0080 %				
				20 kHz to 100 kHz	0,020 %				
				100 kHz to 1 MHz	0,10 %				
				1 MHz to 10 MHz	5,0 %				
		1 MΩ to 10 MΩ		20 Hz to 50 Hz	0,030 %				
				50 Hz to 20 kHz	0,015 %				
				20 kHz to 1 MHz	0,30 %				
		10 MΩ to 100 MΩ		20 Hz to 50 Hz	0,070 %				
				50 Hz to 5 kHz	0,030 %				
				5 kHz to 10 kHz	0,050 %				
				10 kHz to 100 kHz	0,30 %				
21	Secondary component of impedance Z expressed as phase angle φ / AC resistance meters	-π	to	+π	1 mΩ to 10 mΩ 20 Hz to 200 Hz	3.5 μrad	Comparison with a standard using impedance bridge	611-MP-C040 611-MP-C041 611-MP-C099	7



The Appendix is an integral part of  
Certificate of Accreditation No.: 401/2024 of 15/08/2024

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
				200 Hz to 1 kHz 1 kHz to 100 kHz	17·f <sup>4</sup> µrad 17·f µrad			
				10 mΩ to 100 mΩ 100 mΩ to 500 mΩ	20 Hz to 1 kHz 1 kHz to 100 kHz 20 Hz to 1 kHz 1 kHz to 1 MHz	2,5 µrad 4·f µrad 1,5 µrad 1,5·f µrad		
				500 mΩ to 1 Ω 1 Ω to 1 kΩ	20 Hz to 1 kHz 1 kHz to 1 MHz 20 Hz to 1 kHz 1 kHz to 1MHz	1,0 µrad 0,9·f µrad 0,7 µrad 0,7·f µrad		
				1 kΩ to 10 kΩ 10 kΩ to 100 kΩ	20 Hz to 1 kHz 1 kHz to 1 MHz 20 Hz to 1 kHz 1 kHz to 1 MHz	0,9 µrad 0,9·f µrad 1,2 µrad 1,2·f µrad		
				100 kΩ to 10 MΩ	50 Hz to 200 Hz 200 Hz to 5 kHz 5 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 100 kHz	50 µrad 8,0 µrad 20 µrad 30 µrad 1,2·f µrad		
22	Inductance / Inductance meters	1 μH	to	10 mH 100 mH 10 μH	1 kHz 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 100 kHz	0.15 μH 1.5 μH 0.10 % 0.050 % 0.020 %	Comparison with a standard using impedance bridge	611-MP-C099 611-MP-C041



11\_01-P508b K-20240701

-3-

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
				100 kHz to 1 MHz	0.050 %			
		10 µH	to	100 µH	50 Hz to 200 Hz	0.10 %		
					200 Hz to 1 kHz	0.015 %		
					1 kHz to 20 kHz	0.020 %		
					20 kHz to 100 kHz	0.015 %		
					100 kHz to 1 MHz	0.050 %		
		100 µH	to	1 mH	20 Hz to 50 Hz	0.30 %		
					50 Hz to 1 kHz	0.015 %		
					1 kHz to 100 kHz	0.010 %		
					100 kHz to 1 MHz	0.050 %		
		1 mH	to	10 mH	20 Hz to 50 Hz	0.30 %		
					50 Hz to 200 Hz	0.015 %		
					200 Hz to 1 kHz	0.0080 %		
					1 kHz	0.0050 %		
					1 kHz to 20 kHz	0.0080 %		
					20 kHz to 100 kHz	0.015 %		
					100 kHz to 1 MHz	0.050 %		
		10 mH	to	100 mH	20 Hz to 50 Hz	0.30 %		
					50 Hz to 1 kHz	0.0080 %		
					1 kHz	0.0050 %		
					1 kHz to 20 kHz	0.015 %		
					20 kHz to 100 kHz	0.030 %		
					100 kHz to 1 MHz	0.050 %		
		100 mH	to	1 H	50 Hz to 1 kHz	0.0080 %		
					1 kHz	0.0050 %		
					1 kHz to 10 kHz	0.020 %		
					10 kHz to 100 kHz	0.050 %		



11\_01\_P508b K-20240701

-3-

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
		1 H	to	10 H	20 Hz to 50 Hz 50 Hz to 1 kHz 1 kHz to 1 MHz	0.30 % 0.020 % 0.050 %		
		10 H	to	100 H	20 Hz to 50 Hz 50 Hz to 1 kHz 1 kHz to 100 kHz	0.30 % 0.030 % 0.050 %		
		100 H	to	1 kH	20 Hz to 50 Hz 50 Hz to 10 kHz	0.30 % 0.10 %		
23	Digital electrical inspection equipment and associated standards / insulation resistance	10 kΩ	to	10 MΩ	0.0012 % 0.0020 % 0.010 % 0.025 % 0.050 % $0.002 \cdot R^2 + 0.001 \cdot R$ <sup>Note 5</sup> 0.0017 % 0.0030 % 0.010 %	Indirect measurement using a voltage calibrator, reference resistor and voltmeter	131-MP-C006	1.7
		10 MΩ	to	100 MΩ				
		100 MΩ	to	1 GΩ				
		1 GΩ	to	10 GΩ				
		10 GΩ	to	100 GΩ				
		100 GΩ	to	1 TΩ				
				100 MΩ				
				1 GΩ				
				10 GΩ				
		10 kΩ	to	1 GΩ	0.020 %	Direct generation of resistance with a calibrator or inspection		



11\_01-P508b K-20240701

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
	Digital electrical inspection equipment and associated standards / protective loop, network impedance and ground loop	1 GΩ	to	10 GΩ	instruments, resistance decade or resistance reference standards	1.0 %	Direct generation of resistance with a calibrator of inspection instruments, resistance decade or resistance reference standards	
		10 GΩ	to	100 GΩ		1.5 %		
		100 GΩ	to	1 TΩ		2.5 %		
		1 TΩ	to	10 TΩ		4.0 %		
		100 MΩ				0.0030 %		
		1 GΩ				0.0050 %		
		10 GΩ				0.020 %		
		100 GΩ				0.10 %		
		1 TΩ				0.50 %		
		25 mΩ				5.0 mΩ		
		50 mΩ				5.0 mΩ		
		100 mΩ				5.0 mΩ		
		330 mΩ				7.0 mΩ		



11\_01-P508b K-20240701  
-3-

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
		500 mΩ			8.0 mΩ			
		1 Ω			10 mΩ			
		1.8 Ω			18 mΩ			
		5 Ω			30 mΩ			
		10 Ω			60 mΩ			
		18 Ω			100 mΩ			
		50 Ω			300 mΩ			
		100 Ω			500 mΩ			
		180 Ω			1.0 Ω			
		500 Ω			2.5 Ω			
		1 kΩ			5.0 Ω			
		1.8 kΩ			10 Ω			
		3 mA	to	3 A	50 Hz	0.20 %	Simulation of a residual current device and direct measurement of tripping current by a calibrator of inspection instruments	
		3 mA	to	3 A	50 Hz	1.0 %		
	Digital electrical inspection equipment and associated standards / RCD tripping current	10 ms	to	5 s			Simulation of a residual current device and direct measurement of tripping time by a calibrator of inspection instruments	
		11_01-P508b	K-20240701	50 Hz		0.05 ms		



**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
	Digital electrical inspection equipment and associated standards / passive leak current	10 ms	to	5 s	50 Hz	0.020 % + 0.25 ms	Direct comparison with a calibrator of inspection instruments	
	Digital electrical inspection equipment and associated standards / differential leak current	0.1 mA	to	30 mA	50 Hz	0.30 % + 2 µA		
	Digital electrical inspection equipment and associated standards / differential leak current	0.1 mA	to	30 mA	50 Hz	0.30 % + 2 µA		
	Digital electrical inspection equipment and associated standards / active leak current	0.1 mA	to	30 mA	50 Hz	0.30 % + 2 µA		
	Digital electrical inspection equipment and associated standards / active leak current	0.1 mA	to	30 mA	50 Hz	0.30 % + 1 µA		
	Digital electrical inspection equipment and associated standards / DC voltage	0 V	to	10 kV	DC + 20 Hz to 400 Hz	0.20 %		
						0.30 % + 5 V	Direct measurement of high voltage with a calibrator of inspection instruments or a voltmeter and voltage probe	



11\_01-P508b K-20240701

-3-

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

Český metrologický institut  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
	Digital electrical inspection equipment and associated standards / AC voltage	0 V	to	10 kV	50 Hz	0.50 % + 5 V	Current simulation with a calibrator and current coil	
	Digital electrical inspection equipment and associated standards / DC current	1 A	to	1 kA	Current clamp	0.80 %		
	Digital electrical inspection equipment and associated standards / AC current	1 A	to	1 kA	Current clamp	0.50 %		



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit	max.	unit					
24	Oscilloscope vertical deflection coefficient	-222 V	to	222 V		0 Hz	0.025 % + 25 µV	Direct voltage generation with an oscilloscope calibrator	113-MP-C008	1
		1 mV	to	21 mV		10 Hz to 10 kHz	0.1 % + 15 µV			
		21 mV	to	556 mV			0.1 % + 1 µV			
		556 mV	to	210 V			0.05 % + 1 µV			
25	Oscilloscope time base	1 ns	to	50 s			0.25 · 10 <sup>-6</sup>	Direct generation by an oscilloscope calibrator	113-MP-C008	1
26	Oscilloscope bandwidth	0.1 Hz	to	300 MHz		drop -3 dB	4 %	Measurement using oscilloscope calibrator	113-MP-C008	1
		300 MHz	to	550 MHz			5 %			
		550 MHz	to	1.1 GHz			7 %			
		1.1 GHz	to	3.2 GHz			8 %			
	Relative decrease of oscilloscope frequency response	-6 dB	to	6 dB		0.1 Hz to 300 MHz	0.18 dB			
						300 MHz to 550 MHz	0.22 dB			
						550 MHz to 1.1 GHz	0.31 dB			
						1.1 GHz to 3.2 GHz	0.35 dB			
27	Oscilloscope rise time	300 ps	to	1 s				Calibrator signal measurement with an oscilloscope and correction	113-MP-C008	1
							12 ps			



11-01-P508b K-20240701

-3-

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion	
		min.	unit	max.	unit						
28	Input resistance / oscilloscopes, counters, etc.	800 kΩ 40 Ω	to	1.2 MΩ 90 Ω			0.1 % 0.1 %	Measurement using an oscilloscope calibrator	113-MP-C008	1	
29	Voltage/oscilloscopic probe split ratio	0.9 : 1	to	1,100 : 1		To 222 V	0 Hz to 10 kHz	0.5 %	Direct voltage generation with an oscilloscope calibrator	113-MP-C008	1
30*	RF power calibration factor / power sensors	0.05	to	1.1		0 GHz to 1 GHz 1 GHz to 18 GHz 18 GHz to 40 GHz 40 GHz to 50 GHz	0.9 % 1.5 % 2.0 % 3.0 %	Direct comparison of reference and calibrated meter reading	113-MP-C014	1	
	RF power level <i>L</i> /level meters,	44 dB(mW) 20 dB(mW) 10 dB(mW) -10 dB(mW)	to	55 dB(mW) 44 dB(mW) 20 dB(mW) 10 dB(mW)		9 kHz to 2.5 GHz 9 kHz to 6 GHz 6 GHz to 18 GHz 9 kHz to 1 GHz 1 GHz to 10 GHz 10 GHz to 18 GHz 18 GHz to 40 GHz 40 GHz to 50 GHz 9 kHz to 10 GHz 10 GHz to 18 GHz 18 GHz to 40 GHz 40 GHz to 50 GHz	0.12 dB 0.09 dB 0.14 dB 0.06 dB 0.07 dB 0.1 dB 0.13 dB 0.15 dB 0.05 dB 0.09 dB 0.13 dB 0.15 dB				



11\_01-P508b K-20240701  
-3-

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion	
		min.	unit						
		-30 dB(mW)	to	-10 dB(mW)	9 kHz to 1 GHz 1 GHz to 10 GHz 10 GHz to 18 GHz 18 GHz to 40 GHz 40 GHz to 50 GHz 9 kHz to 1 GHz 1 GHz to 10 GHz 10 GHz to 18 GHz 18 GHz to 40 GHz 40 GHz to 50 GHz 0.1 MHz to 1 GHz 1 GHz to 10 GHz 10 GHz to 18 GHz 18 GHz to 26.5 GHz	0.06 dB 0.07 dB 0.1 dB 0.13 dB 0.15 dB 0.07 dB 0.09 dB 0.1 dB -0.001·(L+30)+ 0.13 dB -0.001·(L+30)+ 0.15 dB -0.001·(L+60)+ 0.07 dB -0.001·(L+60)+ 0.07 dB -0.0015·(L+60)+ 0.11 dB -0.002·(L+60)+ 0.15 dB			
	RF voltage at a defined location on the coaxial line/probes, transducers	1 mV	to	2 V	0 GHz to 2 GHz	1 % 1.7 %	Power measurements and recalculations with vector corrections		
	RF current at a defined location on the coaxial line/probes, transducers	20 µA	to	40 mA	0 GHz to 2 GHz	1 % 1.7 %			
	ΔL level difference of power, voltage and current/amplifiers, attenuators, level meters, signal sources	40 mA	to	2 A	9 kHz to 1 MHz	0,0006ΔL + 0.02 dB	Power ratio measurement		
		0 dB	to	65 dB	 11_01-P508b K-20240701				

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit	max.	unit					
	and other test equipment	65 dB	to	75 dB			0.14 dB			
		75 dB	to	85 dB			0.21 dB			
		0 dB	to	85 dB		1 MHz to 13.2 GHz	0,0006ΔL + 0.02 dB			
		85 dB	to	95 dB			0.14 dB			
		95 dB	to	105 dB			0.21 dB			
		105 dB	to	110 dB			0.56 dB			
		0 dB	to	75 dB		13.2 GHz to 26.5 GHz	0,0006ΔL + 0.02 dB			
		75 dB	to	85 dB			0.14 dB			
		85 dB	to	95 dB			0.21 dB			
		95 dB	to	100 dB			0.56 dB			
		0 dB	to	40 dB		26.5 GHz to 50 GHz	0.1 dB			
31	Voltages / equipment used mainly in testing (EMC)	0.5 mV	to	1,000 V		0 Hz	0.1 %	Measurement by a reference standard multimeter	113-MP-C014	1
	Voltage/ESD simulators	1 mV	to	100 V		10 Hz to 100 kHz	0.15 % +3 µV			
		100 V	to	500 V		10 Hz to 100 kHz	0.32 %			
		500 V	to	35 kV		0 Hz	2 % +5 V	Direct measurement by a standard HV voltmeter	113-MP-C017	
32	Current / equipment for testing (EMC)	0.1 mA	to	1 A		0 Hz	0.1 %	Measurement by a reference standard multimeter	113-MP-C014	1
	Current / excitation of frame antenna	1 A	to	40 A		40 Hz to 60 Hz	1.5 %	Measurement using a shunt	113-MP-C017	
		40 A	to	400 A		40 Hz to 60 Hz	1.8 %			



-3- 11\_01-P508b K-20240701

The Appendix is an integral part of  
Certificate of Accreditation No.: 401/2024 of 15/08/2024

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion	
		min.	unit	max.	unit						
	Current / current probes, shunts	50 mA	to	1.9999 A		10 Hz to 5 kHz	0.5 %				
33	Transfer impedance / current probes	-65 dB ( $\Omega$ )	to	+35 dB ( $\Omega$ )		10 Hz to 100 MHz 100 MHz to 300 MHz 300 MHz to 400 MHz	0.2 dB 0.35 dB 0.5 dB	Input current generation and output voltage measurement	113-MP-C017	1	
34	Amplitude modulation depth $m$ / signal sources, modulation meters	5 %	to	99 %		fc: 100 kHz to 10 MHz	fmod: 50 Hz to 10 kHz	0.0075· $m$	Measurement by ref. modulation analyzer, direct comparison with a reference standard	113-MP-C014	
		5 %	to	20 %		fc: 10 MHz to 3 GHz	fmod: 50 Hz to 100 kHz	0.025· $m$ 0.005· $m$			
35	Impedance/ coupling networks	3 $\Omega$	to	200 $\Omega$		9 kHz to 400 MHz	module phase	6 % 4°	Measurement by a vector circuit analyzer and recalculations	113-MP-C017	1
36	Reflection coefficient modulus $r$ / equipment used mainly in testing (EMC)	0	to	1		BNC connector	9 kHz to 1 GHz 1 GHz to 2 GHz 2 GHz to 3 GHz	0.015 $r^2$ + 0.01 0.025 $r^2$ + 0.02 0.035 $r^2$ + 0.03	Measurement by a vector circuit analyzer	113-MP-C017	1



**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit	max.	unit					
37	Current/pulse generators, ESD simulators	1 A	to	130 A		N connector	0.07 $r^2$ + 0.015	Measurements with directional bridges or taps	113-MP-C017	1
						5 kHz to 2 GHz	0.08 $r^2$ + 0.02			
						2 GHz to 8 GHz	0.12 $r^2$ + 0.03			
38	Short circuit current, peak value / pulse generators	1 A	to	3 kA		Rise time > 0.2 $\mu$ s	4 %	Measurement with an oscilloscope with an ESD target	113-MP-C017	1
39	No-load voltage, peak value / pulse generators	20 V	to	8 kV		Rise time > 0.2 $\mu$ s	3.5 %	Measurement using an oscilloscope with a HV probe	113-MP-C017	1
40	Voltage to the load 2 $\Omega$ , 10 $\Omega$ , 20 $\Omega$ , 50 $\Omega$ , peak value / pulse generators	10 V	to	1 kV		Rise time > 0.2 $\mu$ s	3.8 %	Measurement using an oscilloscope with a HV differential probe	113-MP-C017	1
41	Voltage, peak value/pulse generators EFT/burst, US defectoscopes,	10 V	to	4 kV		To 50 $\Omega$ load	3 %	Measurement with an oscilloscope with a HV divider or with attenuation cells	113-MP-C017	1
						To 1 k $\Omega$ load	4 %			



The Appendix is an integral part of  
Certificate of Accreditation No.: 401/2024 of 15/08/2024

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit	max.	unit					
42	Pulse area/pulse generators for EMI receiver calibration, ČSN EN 55016-1-1 ed.4							Measurement with an oscilloscope with attenuation cells and subsequent calculations and corrections	113-MP-C017	1
		0.1 µVs	to	30 µVs		9 kHz to 150 kHz	2.8 %			
		0.01 µVs	to	1 µVs		150 kHz to 30 MHz	2.8 %			
		1 nVs	to	100 nVs		30 MHz to 1 GHz	3 %			



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
43	AC current ratio / Measuring current transformers	0.1	to	5,000	05 A to 5,000 A/ 5 A and 1 A	50 Hz	0.002%	Comparison with a measuring transformer
	Current phase shift / Measuring current transformers	-600 '	to	600 '	0 5 A to 5,000 A/ 5 A and 1 A	50 Hz	0.07'	Comparison with a measuring transformer
44*	AC voltage ratio / Measuring current transformers	50 100 220		5 kV /100 V 10 kV /100 V 22 kV/100 V	50 Hz	0.006 % 0.006 % 0.006 %	Comparison with a measuring transformer	817-MP-C701
			0.4		100 V to 400 kV/5 V to 250 V	50 Hz	0.007 %	
	Measuring voltage transformers / Voltage phase shift	-600 '	to	600 '	5 kV; 10 kV and 22 kV/100 V	50 Hz	0.21'	Comparison with a measuring transformer
					100 V to 400 kV/5 V to 250 V	50 HZ	0.24'	Comparison with a HV divider
45*	Rogowski coils / AC current	0 A	to	10 kA	1 thread		021 %	Comparison with a measuring current transformer
								817-MP-C705
								12



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand		Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min.	unit	max.	unit				
46	Antenna factor / antennas	-10 dB/m	to	+60 dB/m	3 m distance	30 MHz to 100 MHz 100 MHz to 300 MHz 300 MHz to 1 GHz 1 GHz to 18,000 MHz	2.8 dB 2.8 dB 3.0 dB 2.0 dB	Three antenna method	851-MP-C004, chap. 5.1 (ANSI C63.5 chap. 5)
				10 m distance	30 MHz to 100 MHz 100 MHz to 300 MHz 300 MHz to 1 GHz 1 GHz to 18 GHz	2.2 dB 1.5 dB 1.4 dB 2.0 dB			
				3 m distance	30 MHz to 100 MHz 100 MHz to 300 MHz 300 MHz to 1 GHz 1 GHz to 18 GHz	3.3 dB 3.3 dB 3.5 dB 2.2 dB			
				10 m distance	30 MHz to 100 MHz 100 MHz to 300 MHz 300 MHz to 1 GHz 1 GHz to 18 GHz	2.5 dB 1.7 dB 1.7 dB 2.2 dB			
47	Antenna factor / antennas	-10 dB/m	to	+60 dB/m				Substitution method	851-MP-C004, chap. 5.2 (ANSI C63.5 chap. 6)
48	Antenna factor / antennas	-10 dB/m	to	+60 dB/m	1 m distance	30 MHz to 1 GHz 1 GHz to 18 GHz	3.5 dB 3.1 dB	Measurement of transmission of two identical antennas	851-MP-C004, chap. 5.4 (SAE ARP 958, Rev.D, chap 3 and 4)



11\_01\_P508b K-20240701

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand		Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location			
		min.	unit	max.	unit							
49	Antenna factor / frame antennas	-10	dB/m	to	+60	dB/m	1 m distance	10 kHz to 30 MHz	2.2 dB	Measurement of the magnetic field intensity of a standard antenna	851-MP-C004, chap. 5.5 (SAE ARP 958, Rev.D, chap. 6 and 7)	13
50	Antenna factor / rod antennas	-10	dB/m	to	+60	dB/m		9 kHz to 30 MHz	2.1 dB	Adaptive circuit measurement	851-MP-C004, chap. 5.3 (ČSN EN 55016-1-4)	13

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

<sup>4</sup> f is the frequency value in kHz

<sup>5</sup> R is the resistance value in TΩ



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

Český metrologický institut  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

CMC for the field of measured quantity: Magnetic quantities

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min.	unit					
1	Magnetic induction / Reference magnets	0.3 mT	to	2 T	(0.21 to 1) %	Comparison with a standard teslameter	817-MP-C607	12

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

**CMC for the field of measured quantity: Optical quantities**

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion	
		min.	unit						
1	Spectral reflection coefficient R  Spectrophotometers	0 %	to	100 %	8°/t, 8°/d 380 nm to 460 nm 465 nm to 780 nm 0°/45° 380 nm to 780 nm	(0.1 + 0.010·R) % (abs.) (0.12 + 0.008·R)% (abs.) (0.17 + 0.014·R)% (abs.)	Comparison with a colorimetric standard	818-MP-C802	
					8°/t, 8°/d 380 nm to 460 nm 465 nm to 780 nm 0°/45° 380 nm to 780 nm	(0.1 + 0.010·R)% (abs.) (0.12 + 0.008·R)% (abs.) (0.17 + 0.014·R)% (abs.)			
2	Surface colour, colorimetric coordinates  Spectrophotometers, colorimeters	L* a* b*	2 -110 -110	to to to	99 110 110	8°/t, 8°/d, 0°/45°	0.35(abs.) 0.25(abs.) 0.25(abs.)	Comparison with a colorimetric standard	818-MP-C802
	Colorimetric standard	L* a* b*	2 -110 -110	to to to	99 110 110	8°/t, 8°/d, 0°/45°	0.40(abs.) 0.30(abs.) 0.30(abs.)	Measurement by a reference spectrophotometer	12



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion	
		min.	unit						
3	Surface colour, colorimetric coordinates						Comparison with a colorimetric standard	818-MP-C802	
	Spectrophotometers, colorimeters								
	Y	0.3	to	100	8°/t, 8°/d, 0°/45	0.25(abs.)			
	x	0.002	to	0.7		0.0005(abs.)	Measurement by a reference spectrophotometer		
	y	0.002	to	0.8		0.0005(abs.)			
	Colorimetric standard								
4	Surface colour, colorimetric coordinates						Comparison with a colorimetric standard	818-MP-C802	
	Spectrophotometers, colorimeters								
	L	2	to	99	8°/t, 8°/d, 0°/45	0.35(abs.)			
	u'	0.002	to	0.6		0.0005(abs.)	Measurement by a reference spectrophotometer		
	v'	0.002	to	0.6		0.0005(abs.)			
	Colorimetric standard								
	L	2	to	99	8°/t, 8°/d, 0°/45	0.40(abs.)	Measurement by a reference spectrophotometer	12	
	u'	0.002	to	0.6					



**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min. unit	max. unit					
	v'	0.002	to	0.6	0.0008(abs.)			
5	Gloss / Gloss meters, gloss standards	0.1 GU	to	150 GU	20°, 60°, 85°	1.8 GU	ISO 2813	818-MP-C808
6*	Illumination / Light sources	5 lx	to	50,000 lx		2.2 %	Measurement by a reference photometer	818-MP-C801
	Illumination / Luxmeters	1 lx	to	50,000 lx		0.8 %	Comparison with a reference photometer	818-MP-C811
7*	Irradiance / colorimetric boxes	0.1 $\mu\text{W}\cdot\text{cm}^{-2}$	to	1,000 $\text{mW}\cdot\text{cm}^{-2}$	230 nm to 470 nm	5 %	Comparison with a reference photometer	818-MP-C801
8*	Replacement colour temperature / colorimetric (light) boxes	2,000 K	to	10,000 K		30K	Measurement by a reference spectrophotometer	818-MP-C801
9*	Perpendicular spectral transmission / Transmission spectrophotometers, standard filters	0.1	to	1	(200 ≤ λ < 380) nm	0.0012(abs.)	Comparison with a reference photometer	818-MP-C810
		0.001	to	0.1	(200 ≤ λ < 380) nm	0.0001(abs.)		
		0.6	to	1	(380 ≤ λ < 1,000) nm	0.0008(abs.)		
		0.3	to	0.6	(380 ≤ λ < 400) nm	0.0041(abs.)		
		0.3	to	0.6	(400 ≤ λ < 700) nm	0.00056(abs.)		
		0.3	to	0.6	(700 ≤ λ ≤ 1,000) nm	0.0017(abs.)		
		0.02	to	0.3	(380 ≤ λ < 400) nm	0.0015(abs.)		
		0.02	to	0.3	(400 ≤ λ ≤ 1,000) nm	0.00022(abs.)		
		0.001	to	0.02	(380 ≤ λ < 400) nm	0.0002(abs.)		



11\_01-P508b K-20240701

-3-

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min. unit	max. unit					
		0.001	to	0.02	(400 ≤ λ ≤ 1,000) nm	0.00008(abs.)		
10*	Perpendicular spectral absorbance / Transmission spectrophotometers, standard filters							
		0.00	to	1.00 -	(200 ≤ λ < 380) nm	0.0005 to 0.0052(abs.)	Comparison with a reference photometer	818-MP-C810
		1.00	to	3.00	(200 ≤ λ < 380) nm	0.0004 to 0.046(abs.)		
		0.00	to	0.22	(380 ≤ λ < 1,000) nm	0.0003 to 0.0006(abs.)		
		0.22	to	0.52	(380 ≤ λ < 400) nm	0.0030 to 0.0060(abs.)		
		0.22	to	0.52	(400 ≤ λ < 700) nm	0.0004 to 0.0008(abs.)		
		0.22	to	0.52	(700 ≤ λ ≤ 1,000) nm	0.0012 to 0.0025(abs.)		
		0.52	to	1.70	(380 ≤ λ < 400) nm	0.0022 to 0.034(abs.)		
		0.52	to	1.70	(400 ≤ λ ≤ 1,000) nm	0.0003 to 0.0048(abs.)		
		1.70	to	3.00	(380 ≤ λ < 400) nm	0.0044 to 0.097(abs.)		
		1.70	to	3.00	(400 ≤ λ ≤ 1,000) nm	0.0017 to 0.036(abs.)		
11*	Wavelength λ / Transmission spectrophotometers	200 nm	to	1,700 nm		0.2 nm	Comparison with a standard filter	818-MP-C810
	Wavelength λ / Standard filters	200 nm	to	1,700 nm		0.2 nm	Measurement by a reference spectrophotometer	818-MP-C810
12	Transmission optical density / Optical densitometers	0	to	4.5		0.0038(abs.)	Comparison with an optical density standard	818-MP-C812
	Transmission optical density D / Optical density standards	0	to	4.5		0.0038(abs.)	Measurement by a reference densitometer	818-MP-C812



11\_01-P508b K-20240701

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion		
		min.	unit							
13	Optical power / Optical radiometers, fiber power meters	$1 \cdot 10^{-9}$	V	to	10 W	800 nm to 920 nm 920 nm to 960 nm 960 nm to 1,000 nm 1,000 nm to 1,580 nm 1,580 nm to 1,650 nm	0.40% 0.50% 0.45% 0.47% 0.60%	Comparison with a reference detector <sup>4</sup>	818-MP-C813	12
14	Optical attenuation / Attenuators	0 dB		to	65 dB	800 nm to 1,650 nm	0.052 dB (abs.)	Comparison with a reference attenuators	818-MP-C813	12
15	Linearity / Fiber power meters	0 dB		to	0.5 dB	Power level 0 dBm to -65 dBm    800 nm to 1,650 nm	0.004 dB	Comparison with a reference detector <sup>4</sup>	818-MP-C813	12
16	Wavelength / Fiber spectrum analyzers	800 nm		to	1,650 nm		$4 \cdot 10^{-9}$ nm	Comparison with a reference radiation source, spectrometer / wavemeter	818-MP-C813	12
17	Luminous intensity / Standard light sources	1 cd		to	20,000 cd		0.8 %	Measurement by a reference photometer	818-MP-C811	12
18	Brightness / Brightness meters	$1 \text{ cd m}^{-2}$		to	30,000 $\text{cd m}^{-2}$		0.94 %	Primary realization of brightness / Comparison with a reference brightness meter	818-MP-C805	12
19	Luminous flux / Standard light sources	10 lm		to	20,000 lm		1.0 %	Comparison with a standard light source in the integration sphere or on a reference goniophotometer	818-MP-C807	12



**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min. unit	max. unit					
20	Spectral irradiance / Spectroradiometers	0.65 mW m <sup>-2</sup> nm <sup>-1</sup>	to 300 mW m <sup>-2</sup> nm <sup>-1</sup>	300 nm to 400 nm 400 nm to 1,700 nm 1,700 nm to 2,500 nm	3.4 % 3.0 % 4.5 %	Comparison with a standard source of optical radiation	818-MP-C806	12
21	Spectral radiance / Spectroradiometers	1.12 mW sr <sup>-1</sup> m <sup>-2</sup> nm <sup>-1</sup>	to 10.00 mW sr <sup>-1</sup> m <sup>-2</sup> nm <sup>-1</sup>	370 nm to 800 nm 800 nm to 1,000 nm 1,000 nm to 1,700 nm	3.2 % 3.6 % 3.0 %	Comparison with a standard source and standard of spectral reflectance	818-MP-C806	12
22	Wavelength $\lambda$ / spectroradiometer	365 nm	to 923 nm		0.4 nm	Comparison with a standard source	818-MP-C806	12
23	Spectral integral parameters / Sources of optical radiation Replacement colour temperature T <sub>c</sub> Colour rendering index R <sub>a</sub> Light colour, colorimetric coordinates x y Light colour, colorimetric coordinates u' v'	2,000 K 1 0.002 0.002 0.002 0.002	10,000 K to 100 to 0.7 to 0.8 to 0.6 to 0.6		20 K 1.2 (abs.) 0.0025 (abs.) 0.0020 (abs.) 0.0017 (abs.) 0.0015 (abs.)	Measurement by a reference spectrophotometer	818-MP-C806	12



**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
	Spectral integral parameters / Spectrophotometers							
	Replacement colour temperature $T_c$	2,000 K		10,000 K		20 K		
	Colour rendering index $R_a$	1	to	100		0.7 (abs.)		
	Light colour, colorimetric coordinates	x	0.002	to	0.7		0.0014 (abs.)	
		y	0.002	to	0.8		0.0010 (abs.)	
	Light colour, colorimetric coordinates	u'	0.002	to	0.6		0.0009 (abs.)	
		v'	0.002	to	0.6		0.0006 (abs.)	

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

<sup>4</sup> Expressed in dBm:  $P[\text{dBm}] = 10 \cdot \log(P[\text{W}] / 0,001)$ ; uncertainty  $U[\text{dB}] = 10 \cdot \log(1/(1-U[-]))$ .



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

**CMC for the field of measured quantity: Time and frequency quantities**

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit	max.	unit					
1*	Frequency / signal sources, frequency meters						$1 \cdot 10^{-11}$ $1 \cdot 10^{-11}$ 3 Hz	Measurement by ref. counter, direct comparison with a reference standard alternatively using frequency dividers	113-MP-C007	1, 7
		0.01 Hz	to	3 GHz						1
		3 GHz	to	18 GHz						
2*	Period / signal sources, time interval meters						$1 \cdot 10^{-11}$	Measurement by ref. counter, direct comparison with a reference standard	113-MP-C007	1, 7
		5 ns	to	$10^5$ s						
3*	Time interval / signal sources, time interval meters						$(1.1 \cdot 10^{-9} + 1 \cdot 10^{-11} t)$ s $(10 \cdot 10^{-12} + 2 \cdot 10^{-3} t)$ s	Measurement using a counter Measurement by an oscilloscope	113-MP-C007	1, 7
		0 s	to	$10^5$ s						
4*	Simple pulse counting / pulse sources, pulse counters						0	Measurement by ref. counter, direct comparison with a reference standard	113-MP-C007	1, 7
		0	to	$1 \cdot 10^7$						



11\_01-P508b K-20240701

-3-

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion			
		min.	unit	max.	unit								
5	FM frequency deviation / signal sources, modulation meters	0.2 kHz	to	40 kHz		$f_c=250 \text{ kHz to } 10 \text{ MHz}$ $f_{mod}=20 \text{ Hz to } 10 \text{ kHz}$ $\Delta f/f_{mod} > 0.2$	1.5 %	Measurement by ref. modulation meter, direct comparison with a reference standard	113-MP-C007	1			
		0.25 kHz	to	400 kHz		$f_c=10 \text{ MHz to } 6.6 \text{ GHz}$ $f_{mod}=50 \text{ Hz to } 200 \text{ kHz}$ $\Delta f/f_{mod} > 0.2$	1.5 %						
6	Time interval / mechanically operated stopwatch	0.1 s	to	35,999.99 s			16 ms	Direct comparison with a reference standard	113-MP-C013	1, 2, 7			
7	Relative frequency error of time base / stopwatch with LCD	$-1 \cdot 10^{-3}$	to	$1 \cdot 10^{-3}$			$3 \cdot 10^{-7}$	Frequency measurement – capacitive coupling to LCD	113-MP-C013	1, 2, 7			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).



**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

**CMC for the field of measured quantity: Acoustic quantities and mechanical vibration**

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min. unit	max. unit					
1	Sound pressure level / Acoustic calibrators	60 dB	to	134 dB	re $20 \cdot 10^{-6}$ Pa	0.09 dB	Comparison with a standard microphone in relation to frequency and total distortion (ČSN EN 60942)	812-MP-C211
2	Microphone sensitivity / Laboratory standard microphones	-40 dB	to	-24 dB	re $1V \cdot Pa^{-1}$	0.05 dB	Reciprocal calibration method according to nominal sensitivity of 1“ standard microphone (ČSN EN 61094-1, ČSN EN 61094-2)	812-MP-C216

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

**CMC for the field of measured quantity: Physicochemical quantities**

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit	max.	unit					
1*	Moisture of solids / relative moisture meters	4 % to 50 %				Cereals, and oil seeds	0.25 %	Comparison with reference determination	511-MP-C001	6
		0.001 %	to	20 %						
2*	Moisture / Absolute / relative moisture meters	4 %	to	110 %		Plastics	0.005 %	Comparison with reference determination	511-MP-C009	6
3*	Multi-parameter analyzers – relative moisture content of N- substances oil content Zeleny test	4 %	to	50 %		Solids	0.31 %	Comparison with reference determination	511-MP-C003	6
4	Refractive index / refractometers	5 % to 40 %				Efflux cup with a nozzle	0.27 %	Comparison with reference determination	512-MP-C001	6
		10 %	to	80 %						
5	Kinematic viscosity / efflux time	10 ml	to	75 ml		D4 C3 C4 C5 C6 A4	0.29 % 1.3 ml	Comparison with a calibration liquid	616-MP-C003	7
		1.3	to	1.7						



11\_01-P508b K-20240701

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
6	Kinematic viscosity / capillary viscometer constant			0.001 mm <sup>2</sup> ·s <sup>-2</sup> 0.003 mm <sup>2</sup> ·s <sup>-2</sup> 0.01 mm <sup>2</sup> ·s <sup>-2</sup> 0.03 mm <sup>2</sup> ·s <sup>-2</sup> 0.1 mm <sup>2</sup> ·s <sup>-2</sup> 0.3 mm <sup>2</sup> ·s <sup>-2</sup> 1 mm <sup>2</sup> ·s <sup>-2</sup> 3 mm <sup>2</sup> ·s <sup>-2</sup> 10 mm <sup>2</sup> ·s <sup>-2</sup> 30 mm <sup>2</sup> ·s <sup>-2</sup>	1.8·10 <sup>-6</sup> mm <sup>2</sup> ·s <sup>-1</sup> 7.6·10 <sup>-6</sup> mm <sup>2</sup> ·s <sup>-1</sup> 2.9·10 <sup>-5</sup> mm <sup>2</sup> ·s <sup>-1</sup> 9.7·10 <sup>-5</sup> mm <sup>2</sup> ·s <sup>-1</sup> 3.5·10 <sup>-4</sup> mm <sup>2</sup> ·s <sup>-1</sup> 1.1·10 <sup>-3</sup> mm <sup>2</sup> ·s <sup>-1</sup> 3.9·10 <sup>-3</sup> mm <sup>2</sup> ·s <sup>-1</sup> 1.7·10 <sup>-2</sup> mm <sup>2</sup> ·s <sup>-1</sup> 5.0·10 <sup>-2</sup> mm <sup>2</sup> ·s <sup>-1</sup> 1.6·10 <sup>-1</sup> mm <sup>2</sup> ·s <sup>-1</sup>	Comparison with a standard viscometer	616-MP-C002	7
7	Kinematic viscosity / calibration liquids	0.6 mm <sup>2</sup> ·s <sup>-1</sup> 6 mm <sup>2</sup> ·s <sup>-1</sup> 60 mm <sup>2</sup> ·s <sup>-1</sup> 600 mm <sup>2</sup> ·s <sup>-1</sup> 6,000 mm <sup>2</sup> ·s <sup>-1</sup>	to	6 mm <sup>2</sup> ·s <sup>-1</sup> 60 mm <sup>2</sup> ·s <sup>-1</sup> 600 mm <sup>2</sup> ·s <sup>-1</sup> 6,000 mm <sup>2</sup> ·s <sup>-1</sup> 30,000 mm <sup>2</sup> ·s <sup>-1</sup>	0.21 % 0.32 % 0.35 % 0.42 % 0.52 %	Direct measurement using a standard viscometer	616-MP-C002	7
8	Dynamic viscosity / calibration liquids	0.6 mPa·s 6 mPa·s 60 mPa·s 600 mPa·s 6,000 mPa·s	to	6 mPa·s 60 mPa·s 600 mPa·s 6,000 mPa·s 30,000 mPa·s	0.21 % 0.32 % 0.35 % 0.42 % 0.52 %	Direct measurement with standard viscometer and standard densimeter	616-MP-C002	7



11\_01-P508b K-20240701  
-3-

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Český metrologický institut**  
CAB number 2202, CMI Calibration Laboratory  
Okružní 772/31, 638 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit	max.	unit					
9	Dynamic viscosity / rotary viscometers	10 mPa·s	to	150 mPa·s			0.91 %	Comparison with a calibration liquid	616-MP-C003	7
		150 mPa·s	to	400 mPa·s			0.92 %			
		400 mPa·s	to	1,300 mPa·s			1.1 %			
		1,300 mPa·s	to	30,000 mPa·s			1.4 %			
10	Liquid density / glass density meters	620 kg·m <sup>-3</sup>	to	1,850 kg·m <sup>-3</sup>			0.022 kg·m <sup>-3</sup>	Comparison with a standard density meter	616-MP-C004	7
		0 % vol.	to	100 % vol.			0.019 % vol.			
		0 % wt.	to	90 % wt.			0.019 % wt.			
		10 kg·hl <sup>-1</sup>	to	30 kg·hl <sup>-1</sup>			0.019 kg·hl <sup>-1</sup>			
11	Conductometers	0.005 S·m <sup>-1</sup>	to	0.015 S·m <sup>-1</sup>			from 5.6 % to 0.64 %	Comparison with a standard conductometer	616-MP-C005	7
		0.015 S·m <sup>-1</sup>	to	0.15 S·m <sup>-1</sup>			from 0.64 % to 0.19 %			
		0.15 S·m <sup>-1</sup>	to	1.5 S·m <sup>-1</sup>			from 0.19 % to 1.5 %			
		1.5 S·m <sup>-1</sup>	to	10 S·m <sup>-1</sup>			1.5 %			
12	Ethanol concentration / breath alcohol analyzers	0 mg/l	to	1,400 mg/l	Gas phase	from 0.006 mg/l to 0.028 mg/l		By dry gas	114-MP-C004, chap no. 5.3.1	1, 10
13	Ethanol concentration / breath alcohol analyzers	0 mg/l	to	0.480 mg/l	Gas phase	from 0.008 mg/l from 0.013 mg/l		Simulation method		

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

"This document is an appendix to the certificate of accreditation. In case of any discrepancies between the English and Czech versions, the Czech version shall prevail, both for the certificate appendix and the certificate itself."

